

CERTIFICATE OF ANALYSIS FOR

CERTIFIED REFERENCE MATERIAL

OREAS 48

Glacial Till

(District of Rainy River in western Ontario, Canada)



Accredited for compliance with ISO 17034



COA-1822-OREAS48-R0
BUP-70-10-01 Ver:2.0

27-May-2025

Table 1. Certified Value, Uncertainty & Tolerance Intervals for Au by FA in OREAS 48.

Constituent	Certified Value [†]	95 % Expanded Uncertainty		95 % Tolerance Limits	
		Low	High	Low	High
Pb Fire Assay					
Au, Gold (ppb)	3.0	1.6	4.3	2.5	3.4

SI unit equivalents: ppb (parts per billion; $1 \times 10^{-9} \equiv \mu\text{g/kg}$).

[†]The operationally defined measurand meets the requirements of ISO 17034 [9] and all participating laboratories comply with the requirements of ISO 17025 [8].

Table 2. Certified Value, Uncertainty & Tolerance Intervals for other measurands in OREAS 48.

Constituent	Certified Value	95 % Expanded Uncertainty		95 % Tolerance Limits	
		Low	High	Low	High
Aqua Regia Digestion (sample weights 10-50g)					
Au, Gold (ppb)	2.5	1.5	3.6	IND	IND
Borate / Peroxide Fusion ICP					
Ag, Silver (ppm)	< 5	IND	IND	IND	IND
Al, Aluminium (wt.%)	8.02	7.83	8.22	7.86	8.19
As, Arsenic (ppm)	< 4	IND	IND	IND	IND
Ba, Barium (ppm)	431	413	449	422	441
Be, Beryllium (ppm)	0.93	0.68	1.17	0.80	1.05
Ca, Calcium (wt.%)	3.79	3.68	3.90	3.70	3.88
Ce, Cerium (ppm)	35.4	32.8	37.9	34.5	36.2
Co, Cobalt (ppm)	20.0	18.1	21.9	19.1	20.9
Cr, Chromium (ppm)	78	68	88	73	83
Cs, Caesium (ppm)	1.17	0.99	1.35	1.09	1.25
Cu, Copper (ppm)	41.0	35.8	46.2	38.7	43.3
Dy, Dysprosium (ppm)	1.95	1.81	2.10	1.86	2.05
Er, Erbium (ppm)	1.11	1.01	1.21	1.04	1.19
Eu, Europium (ppm)	0.87	0.80	0.94	0.84	0.91
Fe, Iron (wt.%)	4.29	4.17	4.41	4.21	4.37
Ga, Gallium (ppm)	20.6	19.4	21.9	19.8	21.4
Gd, Gadolinium (ppm)	2.48	2.20	2.77	2.32	2.65
Ge, Germanium (ppm)	1.17	0.96	1.37	0.99	1.34
Ho, Holmium (ppm)	0.38	0.34	0.41	0.35	0.40
In, Indium (ppm)	< 0.3	IND	IND	IND	IND
K, Potassium (wt.%)	1.18	1.14	1.23	1.15	1.22
La, Lanthanum (ppm)	16.9	15.4	18.3	16.2	17.5
Li, Lithium (ppm)	19.8	17.6	22.0	18.3	21.3
Lu, Lutetium (ppm)	0.15	0.14	0.17	IND	IND
Mg, Magnesium (wt.%)	1.72	1.68	1.76	1.69	1.75
Mn, Manganese (wt.%)	0.069	0.066	0.072	0.067	0.071
Nb, Niobium (ppm)	3.18	2.66	3.71	2.82	3.54

SI unit equivalents: ppb (parts per billion; $1 \times 10^{-9} \equiv \mu\text{g/kg}$; ppm (parts per million; $1 \times 10^{-6} \equiv \text{mg/kg}$; wt.% (weight per cent) $\equiv \%$ (mass fraction).

Note: intervals may appear asymmetric due to rounding.

IND = indeterminate (due to limited reading resolution of the methods employed. For practical purposes the 95% Expanded Uncertainty can be set between zero and a two times multiple of the upper bound/non-detect limit value).

Table 2 continued.

Constituent	Certified Value	95 % Expanded Uncertainty		95 % Tolerance Limits	
		Low	High	Low	High
Borate / Peroxide Fusion ICP continued					
Nd, Neodymium (ppm)	16.8	15.5	18.0	16.2	17.4
Ni, Nickel (ppm)	41.3	32.7	49.9	36.6	46.0
P, Phosphorus (wt.%)	0.054	0.050	0.058	0.053	0.055
Pb, Lead (ppm)	5.99	3.81	8.17	4.50	7.48
Pr, Praseodymium (ppm)	4.29	3.97	4.62	4.15	4.44
Rb, Rubidium (ppm)	36.6	34.0	39.2	34.7	38.5
Re, Rhenium (ppm)	< 0.01	IND	IND	IND	IND
S, Sulphur (wt.%)	0.047	0.041	0.053	IND	IND
Sc, Scandium (ppm)	14.5	13.4	15.6	IND	IND
Si, Silicon (wt.%)	29.55	29.01	30.08	28.96	30.14
Sm, Samarium (ppm)	2.97	2.62	3.33	2.81	3.14
Sr, Strontium (ppm)	464	440	488	451	478
Tb, Terbium (ppm)	0.35	0.31	0.39	0.32	0.38
Te, Tellurium (ppm)	< 0.5	IND	IND	IND	IND
Th, Thorium (ppm)	2.46	2.24	2.67	2.29	2.63
Ti, Titanium (wt.%)	0.324	0.312	0.337	0.317	0.331
Tl, Thallium (ppm)	0.19	0.15	0.22	IND	IND
Tm, Thulium (ppm)	0.16	0.13	0.18	IND	IND
U, Uranium (ppm)	0.53	0.44	0.63	0.51	0.55
V, Vanadium (ppm)	115	108	121	110	119
W, Tungsten (ppm)	0.78	0.40	1.17	IND	IND
Y, Yttrium (ppm)	10.5	9.7	11.4	10.3	10.8
Yb, Ytterbium (ppm)	1.03	0.89	1.17	0.95	1.11
Zn, Zinc (ppm)	64	57	71	62	66
Zr, Zirconium (ppm)	95	84	106	88	102
4-Acid Digestion					
Ag, Silver (ppm)	0.070	0.056	0.083	0.059	0.080
Al, Aluminium (wt.%)	7.75	7.53	7.98	7.63	7.88
As, Arsenic (ppm)	0.97	0.71	1.23	0.86	1.09
Ba, Barium (ppm)	423	411	435	416	430
Be, Beryllium (ppm)	0.75	0.72	0.79	0.72	0.79
Bi, Bismuth (ppm)	0.12	0.10	0.13	0.11	0.12
Ca, Calcium (wt.%)	3.64	3.55	3.73	3.58	3.70
Cd, Cadmium (ppm)	0.053	0.042	0.065	0.048	0.059
Ce, Cerium (ppm)	33.2	31.1	35.3	32.4	34.0
Co, Cobalt (ppm)	19.4	18.7	20.1	18.9	19.9
Cr, Chromium (ppm)	62	57	67	60	64
Cs, Caesium (ppm)	1.16	1.09	1.23	1.11	1.21

SI unit equivalents: ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg; wt.% (weight per cent) $\equiv$ % (mass fraction).

Note: intervals may appear asymmetric due to rounding.

IND = indeterminate (due to limited reading resolution of the methods employed. For practical purposes the 95% Expanded Uncertainty can be set between zero and a two times multiple of the upper bound/non-detect limit value).

Table 2 continued.

Constituent	Certified Value	95 % Expanded Uncertainty		95 % Tolerance Limits	
		Low	High	Low	High
4-Acid Digestion continued					
Cu, Copper (ppm)	40.0	38.2	41.8	38.8	41.1
Dy, Dysprosium (ppm)	1.90	1.79	2.00	1.83	1.96
Er, Erbium (ppm)	1.06	0.99	1.12	1.01	1.11
Eu, Europium (ppm)	0.84	0.79	0.90	0.81	0.87
Fe, Iron (wt.%)	4.20	4.09	4.31	4.13	4.27
Ga, Gallium (ppm)	19.3	18.5	20.1	18.9	19.8
Gd, Gadolinium (ppm)	2.37	2.21	2.53	2.26	2.49
Hf, Hafnium (ppm)	1.52	1.41	1.64	1.41	1.63
Ho, Holmium (ppm)	0.36	0.33	0.39	0.34	0.37
In, Indium (ppm)	0.032	0.026	0.038	0.029	0.035
K, Potassium (wt.%)	1.16	1.13	1.19	1.15	1.17
La, Lanthanum (ppm)	15.8	14.7	17.0	15.3	16.3
Li, Lithium (ppm)	19.7	18.8	20.6	19.2	20.2
Lu, Lutetium (ppm)	0.14	0.13	0.15	0.13	0.15
Mg, Magnesium (wt.%)	1.65	1.61	1.69	1.62	1.68
Mn, Manganese (wt.%)	0.067	0.065	0.069	0.065	0.068
Mo, Molybdenum (ppm)	0.81	0.70	0.93	0.74	0.89
Na, Sodium (wt.%)	3.04	2.96	3.12	3.00	3.08
Nb, Niobium (ppm)	3.05	2.90	3.20	2.93	3.17
Nd, Neodymium (ppm)	15.6	14.4	16.8	15.0	16.1
Ni, Nickel (ppm)	38.5	37.1	39.8	37.5	39.4
P, Phosphorus (wt.%)	0.053	0.051	0.055	0.052	0.054
Pb, Lead (ppm)	5.40	5.00	5.79	5.06	5.73
Pr, Praseodymium (ppm)	3.96	3.74	4.18	3.83	4.09
Rb, Rubidium (ppm)	33.7	31.5	35.9	32.6	34.9
S, Sulphur (wt.%)	0.050	0.049	0.051	0.048	0.052
Sb, Antimony (ppm)	0.10	0.08	0.12	0.09	0.11
Sc, Scandium (ppm)	14.9	14.3	15.5	14.4	15.4
Sm, Samarium (ppm)	2.85	2.66	3.04	2.75	2.95
Sn, Tin (ppm)	0.60	0.55	0.65	0.56	0.64
Sr, Strontium (ppm)	467	454	479	458	475
Ta, Tantalum (ppm)	0.21	0.19	0.23	0.21	0.22
Tb, Terbium (ppm)	0.33	0.31	0.35	0.31	0.34
Th, Thorium (ppm)	2.42	2.19	2.65	2.31	2.54
Ti, Titanium (wt.%)	0.318	0.309	0.328	0.311	0.326
Tl, Thallium (ppm)	0.18	0.17	0.20	0.18	0.19
Tm, Thulium (ppm)	0.15	0.13	0.16	0.14	0.16
U, Uranium (ppm)	0.48	0.45	0.52	0.45	0.51
V, Vanadium (ppm)	111	108	114	109	113

SI unit equivalents: ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg; wt.% (weight per cent) $\equiv$ % (mass fraction).

Note: intervals may appear asymmetric due to rounding.

IND = indeterminate (due to limited reading resolution of the methods employed).

Table 2 continued.

Constituent	Certified Value	95 % Expanded Uncertainty		95 % Tolerance Limits	
		Low	High	Low	High
4-Acid Digestion continued					
W, Tungsten (ppm)	0.60	0.53	0.68	0.55	0.66
Y, Yttrium (ppm)	10.1	9.6	10.7	9.9	10.4
Yb, Ytterbium (ppm)	0.94	0.87	1.02	0.90	0.99
Zn, Zinc (ppm)	60	57	62	58	61
Zr, Zirconium (ppm)	49.7	46.4	53.0	47.7	51.7
Aqua Regia Digestion					
Ag, Silver (ppm)	0.057	0.049	0.065	0.051	0.063
Al, Aluminium (wt.%)	1.29	1.23	1.34	1.25	1.32
As, Arsenic (ppm)	0.85	0.67	1.04	0.80	0.91
B, Boron (ppm)	< 10	IND	IND	IND	IND
Ba, Barium (ppm)	67	64	70	65	69
Be, Beryllium (ppm)	0.12	0.11	0.13	IND	IND
Bi, Bismuth (ppm)	0.066	0.055	0.078	0.062	0.071
Ca, Calcium (wt.%)	0.905	0.827	0.983	0.876	0.934
Cd, Cadmium (ppm)	0.029	0.023	0.035	0.024	0.034
Ce, Cerium (ppm)	26.8	25.3	28.2	26.0	27.6
Co, Cobalt (ppm)	11.8	11.3	12.2	11.5	12.1
Cr, Chromium (ppm)	36.8	35.1	38.5	35.6	38.0
Cs, Caesium (ppm)	0.73	0.68	0.78	0.70	0.76
Cu, Copper (ppm)	39.7	38.3	41.2	38.6	40.9
Dy, Dysprosium (ppm)	0.96	0.79	1.14	0.92	1.00
Eu, Europium (ppm)	0.43	0.33	0.53	0.41	0.45
Fe, Iron (wt.%)	2.10	1.96	2.23	2.04	2.15
Ga, Gallium (ppm)	4.48	4.21	4.76	4.36	4.61
Gd, Gadolinium (ppm)	1.47	1.20	1.74	1.40	1.55
Hf, Hafnium (ppm)	0.19	0.17	0.21	0.18	0.20
Ho, Holmium (ppm)	0.17	0.14	0.20	IND	IND
In, Indium (ppm)	0.009	0.007	0.012	IND	IND
K, Potassium (wt.%)	0.264	0.252	0.277	0.256	0.272
La, Lanthanum (ppm)	12.6	12.0	13.2	12.2	13.0
Li, Lithium (ppm)	14.9	14.3	15.6	14.6	15.3
Lu, Lutetium (ppm)	0.053	0.043	0.062	0.048	0.057
Mg, Magnesium (wt.%)	0.808	0.778	0.839	0.788	0.828
Mn, Manganese (wt.%)	0.030	0.029	0.031	0.029	0.031
Mo, Molybdenum (ppm)	0.72	0.61	0.82	0.65	0.79
Na, Sodium (wt.%)	0.095	0.088	0.102	0.093	0.097
Nb, Niobium (ppm)	0.16	0.12	0.20	0.14	0.18
Nd, Neodymium (ppm)	11.5	10.4	12.7	11.2	11.9

SI unit equivalents: ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg; wt.% (weight per cent) $\equiv$ % (mass fraction).

Note: intervals may appear asymmetric due to rounding.

IND = indeterminate (due to limited reading resolution of the methods employed. For practical purposes the 95% Expanded Uncertainty can be set between zero and a two times multiple of the upper bound/non-detect limit value).

Table 2 continued.

Constituent	Certified Value	95 % Expanded Uncertainty		95 % Tolerance Limits	
		Low	High	Low	High
Aqua Regia Digestion continued					
Ni, Nickel (ppm)	23.4	22.2	24.6	22.9	24.0
P, Phosphorus (wt.%)	0.051	0.049	0.053	0.050	0.052
Pb, Lead (ppm)	1.50	1.32	1.68	1.41	1.58
Pr, Praseodymium (ppm)	3.05	2.76	3.34	2.99	3.10
Rb, Rubidium (ppm)	13.4	12.8	14.1	13.0	13.9
Re, Rhenium (ppm)	0.000	0.000	0.001	IND	IND
S, Sulphur (wt.%)	0.050	0.050	0.050	IND	IND
Sc, Scandium (ppm)	3.20	2.95	3.46	3.07	3.34
Sm, Samarium (ppm)	1.82	1.63	2.01	1.74	1.90
Sn, Tin (ppm)	0.25	0.21	0.29	0.24	0.26
Sr, Strontium (ppm)	45.4	41.3	49.4	44.3	46.5
Ta, Tantalum (ppm)	< 0.01	IND	IND	IND	IND
Te, Tellurium (ppm)	0.020	0.014	0.026	IND	IND
Th, Thorium (ppm)	2.04	1.88	2.20	1.94	2.14
Ti, Titanium (wt.%)	0.133	0.120	0.145	0.129	0.136
Tl, Thallium (ppm)	0.094	0.082	0.106	0.090	0.098
U, Uranium (ppm)	0.25	0.23	0.28	0.24	0.27
V, Vanadium (ppm)	43.5	41.2	45.9	42.7	44.4
W, Tungsten (ppm)	0.19	0.15	0.22	0.16	0.21
Y, Yttrium (ppm)	4.17	3.79	4.55	4.03	4.31
Yb, Ytterbium (ppm)	0.38	0.32	0.45	0.36	0.41
Zn, Zinc (ppm)	40.6	38.7	42.4	39.4	41.7
Zr, Zirconium (ppm)	4.97	4.51	5.43	4.74	5.19

SI unit equivalents: ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg; wt.% (weight per cent) $\equiv$ % (mass fraction).

Note: intervals may appear asymmetric due to rounding.

IND = indeterminate (due to limited reading resolution of the methods employed. For practical purposes the 95% Expanded Uncertainty can be set between zero and a two times multiple of the upper bound/non-detect limit value).

Table 3. Indicative Values for OREAS 48.

Constituent	Unit	Value	Constituent	Unit	Value	Constituent	Unit	Value
Borate / Peroxide Fusion ICP								
B	ppm	19.7	Mo	ppm	1.72	Sn	ppm	1.80
Bi	ppm	0.15	Na	wt. %	2.79	Ta	ppm	0.35
Cd	ppm	< 0.8	Sb	ppm	0.37			
Hf	ppm	2.76	Se	ppm	< 3			
4-Acid Digestion								
Au	ppm	0.006	Hg	ppm	0.018	Re	ppm	0.001
B	ppm	4.41	Pd	ppb	< 2	Se	ppm	0.11
Ge	ppm	0.079	Pt	ppb	3.17	Te	ppm	0.031
Aqua Regia Digestion								
Er	ppm	0.43	Pd	ppb	2.08	Se	ppm	0.19
Ge	ppm	0.047	Pt	ppb	1.41	Tb	ppm	0.17
Hg	ppm	0.011	Rh	ppm	0.001	Tm	ppm	0.068
Ir	ppm	0.000	Sb	ppm	0.050			
Borate Fusion XRF								
Al ₂ O ₃	wt. %	15.07	K ₂ O	wt. %	1.38	SO ₃	wt. %	0.142
BaO	ppm	533	MgO	wt. %	2.86	SrO	ppm	410
CaO	wt. %	5.11	MnO	wt. %	0.093	TiO ₂	wt. %	0.520
Cr ₂ O ₃	ppm	112	Na ₂ O	wt. %	4.26	V ₂ O ₅	ppm	207
CuO	ppm	55	NiO	ppm	23.3	ZnO	ppm	92
Fe ₂ O ₃	wt. %	6.12	P ₂ O ₅	wt. %	0.122	ZrO ₂	ppm	148
HfO ₂	ppm	15.0	SiO ₂	wt. %	61.77			

SI unit equivalents: ppb (parts per billion; 1×10^{-9}) $\equiv$ $\mu\text{g/kg}$; ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg ; wt. % (weight per cent) $\equiv$ % (mass fraction). Note: the number of significant figures reported is not a reflection of the level of certainty of stated values. They are instead an artefact of ORE's in-house CRM-specific LIMS.

TABLE OF CONTENTS

INTRODUCTION	9
SOURCE MATERIAL.....	9
COMMINUTION AND HOMOGENISATION PROCEDURES	9
PHYSICAL PROPERTIES	10
MINERALOGY	10
ANALYTICAL PROGRAM.....	10
STATISTICAL ANALYSIS	11
Certified Values and their uncertainty intervals	11
Indicative (uncertified) values.....	11
Homogeneity Evaluation	11
PERFORMANCE GATES	12
PARTICIPATING LABORATORIES.....	17
PREPARER AND SUPPLIER.....	18
METROLOGICAL TRACEABILITY	18
COMMUTABILITY	19
INTENDED USE	19
MINIMUM SAMPLE SIZE	19
PERIOD OF VALIDITY & STORAGE INSTRUCTIONS.....	20
INSTRUCTIONS FOR HANDLING & CORRECT USE.....	20
LEGAL NOTICE	21
QMS CERTIFICATION	21
DOCUMENT HISTORY	22
CERTIFYING OFFICER.....	22
REFERENCES	22

LIST OF TABLES

Table 1. Certified Value, Uncertainty & Tolerance Intervals for Au by FA in OREAS 48.	2
Table 2. Certified Value, Uncertainty & Tolerance Intervals for other measurands in OREAS 48.	2
Table 3. Indicative Values for OREAS 48.	7
Table 4. Physical properties of OREAS 48.	10
Table 5. Indicative mineralogy of OREAS 48 based on semi-quantitative XRD analysis...	10
Table 6. Performance Gates for OREAS 48.	13

INTRODUCTION

OREAS reference materials are intended to provide a low-cost method of evaluating and improving the quality of analysis of geological samples. To the geologist they provide a means of implementing quality control in analytical data sets generated in exploration from the grass roots level through to prospect evaluation, and in grade control at mining operations. To the analyst they provide an effective means of calibrating analytical equipment, assessing new techniques and routinely monitoring in-house procedures. OREAS reference materials enable users to successfully achieve process control of these tasks because the observed variance from repeated analysis has its origin almost exclusively in the analytical process rather than the reference material itself. In evaluating laboratory performance with this CRM, the section headed 'Instructions for correct use' should be read carefully.

Table 1 (generated from data supplied by laboratories all accredited to ISO 17025 for Au by fire assay) and Table 2 (generated from data supplied by laboratories mostly accredited to ISO 17025) provide the certified values and their associated 95 % expanded uncertainty and tolerance intervals, Table 3 shows indicative values including major and trace element characterisation, Table 4 provides some indicative physical properties, Table 5 provides indicative mineralogy based on semi-quantitative XRD analysis and Table 6 presents the performance gate intervals for all certified values.

Tabulated results of all analytes together with uncorrected means, medians, standard deviations, relative standard deviations and per cent deviation of lab means from the corrected mean of means (PDM³) are presented in the detailed certification data for this CRM (**OREAS 48-DataPack.1.0.250505_185301.xlsx**).

SOURCE MATERIAL

Certified Reference Material (CRM) OREAS 48 is derived from glacial till collected in the Rainy River District of western Ontario. This till, predominantly of Late Wisconsinan age, is typically classified as an unsorted diamicton. Its composition reflects both the local Precambrian Shield bedrock including granitoid rocks, metavolcanics, and metasedimentary rocks, and Paleozoic sedimentary formations. The material contains a mix of felsic to mafic lithologies with mineral constituents such as quartz, feldspar, amphibole, and biotite, along with occasional sulphide minerals. The till's combined local and long-range glacial provenance makes it a reliable matrix for geochemical exploration, particularly for QAQC applications targeting background concentrations of gold, PGE, REE, lithium, and base metals.

COMMINUTION AND HOMOGENISATION PROCEDURES

The material constituting OREAS 48 was prepared in the following manner:

- 72-hour sterilising heat treatment by drying at 110 °C;
- Multi-stage milling to 98% minus 75 microns;
- Homogenisation using OREAS' novel processing technologies;
- Packaging in 10g and 60g units in laminated foil pouches and 1kg units in plastic wide-mouth jars.

PHYSICAL PROPERTIES

OREAS 48 was tested at ORE Research & Exploration Pty Ltd's onsite facility for various physical properties. Table 4 presents these findings that should be used for informational purposes only.

Table 4. Physical properties of OREAS 48.

Bulk Density (kg/m ³)	Moisture (wt.%)	Munsell Notation [‡]	Munsell Color [‡]
1066	0.15	N7	Light Gray

[‡]The Munsell Rock Color Chart helps geologists and archeologists communicate with colour more effectively by cross-referencing ISCC-NBS colour names with unique Munsell alpha-numeric colour notations for rock colour samples.

MINERALOGY

The semi-quantitative XRD results shown in Table 5 below were undertaken by ALS Metallurgy in Balcatta, Western Australia. The results have been normalised to 100 % and represent the relative proportion of crystalline material. Totals greater or less than 100 % are due to rounding errors. Clay mineral is mainly vermiculite, illite, and smectite. A trace of lowaite and/or Zunyite might be present in the samples.

Table 5. Indicative mineralogy of OREAS 48 based on semi-quantitative XRD analysis.

Mineral / Mineral Group	% (mass ratio)
Clay mineral	1
Chlorite	9
Annite - biotite - phlogopite	6
Muscovite	6
Ca-Na amphibole	25
Clinopyroxene	< 1
Plagioclase	30
K-feldspar	2
Epidote	2
Quartz	17
Calcite	< 1
Siderite	1
Pyrite	< 1
Magnetite	< 1

ANALYTICAL PROGRAM

Twenty commercial analytical laboratories participated in the program to certify the elements reported in Table 1 and 2. The following methods were employed:

- Gold by 15-50g fire assay with ICP-OES (10 laboratories), ICP-MS (4 laboratories) and AAS (4 laboratories) finish;

- Gold by aqua regia digestion (10-50g sample weight) with ICP-OES and/or ICP-MS (9 laboratories) finish and AAS (5 laboratories) finish;
- Lithium borate fusion or sodium peroxide fusion for full ICP-OES and ICP-MS elemental suites (up to 17 laboratories depending on the element; Only two laboratories used sodium peroxide fusion);
- Full ICP-OES and ICP-MS elemental suites by 4-acid ($\text{HNO}_3\text{-HF-HClO}_4\text{-HCl}$) digestion (up to 20 laboratories depending on the element);
- Full ICP-OES and ICP-MS elemental suites by aqua regia digestion (up to 18 laboratories depending on the element).

For the round robin program twelve 1.8 kg test units were taken at predetermined intervals during the bagging stage, immediately following homogenisation and are considered representative of the entire prepared batch. The six samples received by each laboratory were obtained by taking a 120 g subsample from either the *odd* or *even* sampling (lot) intervals to maximise representation.

To confirm homogeneity, the repeatability of trace elements was evaluated using fused bead laser ablation with ICP-MS at Bureau Veritas in Perth, Australia. This technique offers high precision even at low detection levels (see 'Homogeneity Evaluation' section below).

STATISTICAL ANALYSIS

Certified Values and their uncertainty intervals (Tables 1 and 2) have been determined for each analyte following removal of individual, laboratory dataset (batch) and 3SD outliers (single iteration).

For individual outliers within a laboratory batch the z-score test is used in combination with a second method that determines the per cent deviation of the individual value from the batch median. Outliers in general are selected on the basis of z-scores > 2.5 and with per cent deviations (i) > 3 and (ii) more than three times the average absolute per cent deviation for the batch. Each laboratory data set mean is tested for outlying status based on z-score discrimination and rejected if > 2.5 . After individual and laboratory data set (batch) outliers have been eliminated a non-iterative 3 standard deviation filter is applied, with those values lying outside this window also relegated to outlying status. However, while statistics are taken into account, the exercise of a statistician's prerogative plays a significant role in identifying outliers.

95 % Expanded Uncertainty provides a 95 % probability that the true value of the analyte under consideration lies between the upper and lower limits and is calculated according to the method outlined in [5] and [15]. All known or suspected sources of bias have been investigated or taken into account.

Indicative (uncertified) values (Table 3) are present where the number of laboratories reporting a particular analyte is insufficient (< 5) to support certification or where interlaboratory consensus is poor. This data is intended for 'informational purposes' only.

Homogeneity Evaluation

The tolerance limits (ISO 16269:2014 [6]) shown in Tables 1 and 2 were determined using an analysis of precision errors method and are considered a conservative estimate of true homogeneity. The meaning of tolerance limits may be illustrated for copper by 4-acid digestion, where 99 % of the time ($1-\alpha=0.99$) at least 95 % of subsamples ($p=0.95$) will have

concentrations lying between 38.8 and 41.1 ppm. Put more precisely, this means that if the same number of subsamples were taken and analysed in the same manner repeatedly, 99 % of the tolerance intervals so constructed would cover at least 95 % of the total population, and 1% of the tolerance intervals would cover less than 95 % of the total population. **Please note that tolerance limits pertain to the homogeneity of the CRM only and should not be used as control limits for laboratory performance.**

In addition to the precision error method outlined above, homogeneity was also evaluated using a study of repeatability errors of trace elements using fused bead laser ablation with ICP-MS. This involved sending 12 x 10 g samples (one from each of the twelve 1.8 kg test units) to Bureau Veritas in Perth, Australia. This technique offers high precision even at low detection levels. Only analytes present in concentration levels greater than 20 x LLD were evaluated in order to avoid the impact of reading resolution errors in the assessment of repeatability. Of the 50 trace elements available, 32 were deemed appropriate for homogeneity assessment (present in concentrations greater than 20 x LLD). The RSD's for 29 of these 32 elements were under 5 % with 15 under 3 % and 11 under 2 %. Those under 2% RSD included Ce, Sr, Rb, Ti, Mn, Ga, Ba, Y, V, La and Sc. The 3 elements above 5 % RSD included U, Yb and Ta and all of these elements are present in very low concentrations (0.2 to 1.1 ppm) where this degree of repeatability close to LLD is acceptable.

Based on the statistical analysis of the results of the interlaboratory certification program, it can be concluded that OREAS 48 is fit-for-purpose as a certified reference material (see 'Intended Use' below).

PERFORMANCE GATES

The standard deviations (SD's) intervals reported in Table 6 provide an indication of a level of performance that might reasonably be expected from a laboratory being monitored by this CRM in a QA/QC program. They take into account errors attributable to measurement uncertainty and CRM variability. For an effective CRM the contribution of the latter should be negligible in comparison to measurement errors. The Standard Deviation values include all sources of measurement uncertainty: between-lab variance, within-run variance (precision errors) and CRM variability.

In the application of SD's in monitoring performance it is important to note that not all laboratories function at the same level of proficiency and that different methods in use at a particular laboratory have differing levels of precision. Each laboratory has its own inherent SD (for a specific concentration level and analyte-method pair) based on the analytical process and this SD is not directly related to the round robin program (see 'Intended Use' section for more detail). The SD for each analyte's certified value is calculated from the same filtered data set used to determine the certified value, i.e., after removal of all individual, lab dataset (batch) and 3SD outliers (single iteration). These outliers can only be removed after the absolute homogeneity of the CRM has been independently established, i.e., the outliers must be confidently deemed to be analytical rather than arising from inhomogeneity of the CRM. **The standard deviation is then calculated for each analyte from the pooled accepted analyses generated from the certification program.**

Table 6 below shows intervals calculated for two and three standard deviations. As a guide these intervals may be regarded as warning or rejection for multiple 2SD outliers, or rejection for individual 3SD outliers in QC monitoring, although their precise application should be at the discretion of the QC manager concerned (also see 'Intended Use' section below). Westgard Rules extend the basics of single-rule QC monitoring using multi-rules (for more

information visit www.westgard.com/mltirule.htm). A second method utilises a 5% window calculated directly from the certified value.

Standard deviation is also shown in relative percent for one, two and three relative standard deviations (1RSD, 2RSD and 3RSD) to facilitate an appreciation of the magnitude of these numbers and a comparison with the 5% window. Caution should be exercised when concentration levels approach lower limits of detection of the analytical methods employed as performance gates calculated from standard deviations tend to be excessively wide whereas those determined by the 5% method are too narrow. One approach used at commercial laboratories is to set the acceptance criteria at twice the detection level (DL) $\pm 10\%$.

i.e., Certified Value $\pm 10\% \pm 2DL$ [1].

Table 6. Performance Gates for OREAS 48.

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5 % window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
Pb Fire Assay											
Au, ppb	3.0	1.40	0.2	5.8	0.0	7.2	47.29%	94.59%	141.88%	2.8	3.1
Aqua Regia Digestion (sample weights 10-50g)											
Au, ppb	2.5	0.59	1.4	3.7	0.8	4.3	23.16%	46.32%	69.48%	2.4	2.7
Borate / Peroxide Fusion ICP											
Ag, ppm	< 5	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Al, wt.%	8.02	0.193	7.64	8.41	7.45	8.60	2.40%	4.80%	7.20%	7.62	8.43
As, ppm	< 4	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Ba, ppm	431	17	398	464	381	481	3.84%	7.68%	11.52%	410	453
Be, ppm	0.93	0.21	0.52	1.34	0.31	1.54	22.08%	44.16%	66.23%	0.88	0.98
Ca, wt.%	3.79	0.105	3.58	4.00	3.48	4.11	2.77%	5.55%	8.32%	3.60	3.98
Ce, ppm	35.4	2.41	30.5	40.2	28.1	42.6	6.82%	13.63%	20.45%	33.6	37.1
Co, ppm	20.0	1.65	16.7	23.3	15.0	24.9	8.28%	16.55%	24.83%	19.0	21.0
Cr, ppm	78	7.4	63	93	56	100	9.47%	18.93%	28.40%	74	82
Cs, ppm	1.17	0.15	0.87	1.47	0.72	1.62	12.80%	25.60%	38.39%	1.11	1.23
Cu, ppm	41.0	3.45	34.1	47.9	30.7	51.3	8.41%	16.81%	25.22%	39.0	43.1
Dy, ppm	1.95	0.115	1.72	2.18	1.61	2.30	5.90%	11.80%	17.71%	1.86	2.05
Er, ppm	1.11	0.060	0.99	1.23	0.93	1.29	5.37%	10.75%	16.12%	1.06	1.17
Eu, ppm	0.87	0.045	0.78	0.96	0.74	1.01	5.22%	10.43%	15.65%	0.83	0.92
Fe, wt.%	4.29	0.101	4.09	4.49	3.99	4.60	2.36%	4.72%	7.08%	4.08	4.51
Ga, ppm	20.6	1.07	18.5	22.8	17.4	23.9	5.20%	10.40%	15.59%	19.6	21.7
Gd, ppm	2.48	0.177	2.13	2.84	1.95	3.01	7.13%	14.26%	21.38%	2.36	2.61
Ge, ppm	1.17	0.22	0.72	1.61	0.50	1.84	19.13%	38.27%	57.40%	1.11	1.22
Ho, ppm	0.38	0.023	0.33	0.42	0.31	0.44	6.11%	12.22%	18.34%	0.36	0.39
In, ppm	< 0.3	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
K, wt.%	1.18	0.047	1.09	1.28	1.04	1.32	3.99%	7.98%	11.97%	1.12	1.24
La, ppm	16.9	1.59	13.7	20.1	12.1	21.6	9.39%	18.78%	28.17%	16.0	17.7
Li, ppm	19.8	1.56	16.7	22.9	15.1	24.5	7.86%	15.72%	23.58%	18.8	20.8
Lu, ppm	0.15	0.02	0.12	0.19	0.10	0.21	11.10%	22.21%	33.31%	0.15	0.16
Mg, wt. %	1.72	0.044	1.63	1.81	1.59	1.85	2.58%	5.16%	7.74%	1.63	1.81

SI unit equivalents: ppb (parts per billion; 1×10^{-9}) $\equiv$ $\mu\text{g/kg}$; ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg ; wt. % (weight per cent) $\equiv$ % (mass fraction).

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 6 continued.

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5 % window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
Borate / Peroxide Fusion ICP continued											
Mn, wt. %	0.069	0.002	0.064	0.073	0.062	0.076	3.25%	6.50%	9.75%	0.066	0.072
Nb, ppm	3.18	0.51	2.16	4.20	1.65	4.71	16.01%	32.01%	48.02%	3.02	3.34
Nd, ppm	16.8	1.38	14.0	19.6	12.6	20.9	8.23%	16.47%	24.70%	15.9	17.6
Ni, ppm	41.3	5.4	30.5	52.1	25.1	57.5	13.07%	26.14%	39.22%	39.2	43.4
P, wt. %	0.054	0.004	0.047	0.062	0.043	0.065	6.91%	13.81%	20.72%	0.052	0.057
Pb, ppm	5.99	1.61	2.77	9.20	1.17	10.81	26.84%	53.68%	80.52%	5.69	6.29
Pr, ppm	4.29	0.266	3.76	4.82	3.50	5.09	6.19%	12.37%	18.56%	4.08	4.51
Rb, ppm	36.6	2.19	32.2	40.9	30.0	43.1	5.98%	11.96%	17.94%	34.7	38.4
Re, ppm	< 0.01	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
S, wt. %	0.047	0.007	0.032	0.061	0.025	0.069	15.83%	31.66%	47.50%	0.044	0.049
Sc, ppm	14.5	0.87	12.8	16.3	11.9	17.1	5.97%	11.95%	17.92%	13.8	15.2
Si, wt. %	29.55	0.345	28.86	30.24	28.51	30.58	1.17%	2.34%	3.51%	28.07	31.02
Sm, ppm	2.97	0.234	2.51	3.44	2.27	3.68	7.87%	15.74%	23.60%	2.83	3.12
Sr, ppm	464	21	422	506	401	527	4.53%	9.05%	13.58%	441	487
Tb, ppm	0.35	0.022	0.31	0.39	0.29	0.42	6.13%	12.27%	18.40%	0.33	0.37
Te, ppm	< 0.5	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Th, ppm	2.46	0.105	2.25	2.67	2.14	2.78	4.29%	8.58%	12.87%	2.34	2.58
Ti, wt. %	0.324	0.009	0.307	0.341	0.299	0.350	2.64%	5.28%	7.92%	0.308	0.340
Tl, ppm	0.19	0.02	0.15	0.23	0.13	0.25	10.81%	21.62%	32.42%	0.18	0.19
Tm, ppm	0.16	0.02	0.11	0.20	0.09	0.22	13.28%	26.56%	39.85%	0.15	0.16
U, ppm	0.53	0.06	0.41	0.66	0.34	0.72	11.82%	23.65%	35.47%	0.51	0.56
V, ppm	115	6	103	126	98	132	4.94%	9.87%	14.81%	109	121
W, ppm	0.78	0.22	0.34	1.23	0.12	1.45	28.43%	56.87%	85.30%	0.75	0.82
Y, ppm	10.5	0.96	8.6	12.5	7.7	13.4	9.12%	18.25%	27.37%	10.0	11.1
Yb, ppm	1.03	0.11	0.80	1.26	0.69	1.37	11.12%	22.24%	33.36%	0.98	1.08
Zn, ppm	64	7	49	78	42	85	11.31%	22.62%	33.93%	60	67
Zr, ppm	95	8.6	78	112	69	121	9.04%	18.08%	27.12%	90	100
4-Acid Digestion											
Ag, ppm	0.070	0.014	0.042	0.097	0.028	0.111	19.73%	39.46%	59.18%	0.066	0.073
Al, wt. %	7.75	0.228	7.30	8.21	7.07	8.44	2.94%	5.89%	8.83%	7.37	8.14
As, ppm	0.97	0.17	0.63	1.32	0.45	1.49	17.83%	35.65%	53.48%	0.92	1.02
Ba, ppm	423	17	389	457	372	474	4.04%	8.08%	12.11%	402	444
Be, ppm	0.75	0.041	0.67	0.84	0.63	0.88	5.44%	10.87%	16.31%	0.72	0.79
Bi, ppm	0.12	0.011	0.09	0.14	0.08	0.15	9.69%	19.38%	29.06%	0.11	0.12
Ca, wt. %	3.64	0.107	3.43	3.85	3.32	3.96	2.95%	5.89%	8.84%	3.46	3.82
Cd, ppm	0.053	0.008	0.038	0.069	0.030	0.077	14.72%	29.44%	44.16%	0.051	0.056
Ce, ppm	33.2	2.52	28.1	38.2	25.6	40.7	7.58%	15.17%	22.75%	31.5	34.8
Co, ppm	19.4	0.66	18.1	20.7	17.4	21.4	3.42%	6.85%	10.27%	18.4	20.4
Cr, ppm	62	8	46	78	38	86	12.75%	25.50%	38.25%	59	65
Cs, ppm	1.16	0.061	1.04	1.28	0.98	1.34	5.23%	10.46%	15.69%	1.10	1.22
Cu, ppm	40.0	1.34	37.3	42.6	35.9	44.0	3.36%	6.71%	10.07%	38.0	42.0
Dy, ppm	1.90	0.106	1.68	2.11	1.58	2.21	5.57%	11.14%	16.71%	1.80	1.99
Er, ppm	1.06	0.063	0.93	1.18	0.87	1.24	5.96%	11.92%	17.88%	1.00	1.11

SI unit equivalents: ppb (parts per billion; 1×10^{-9}) $\equiv$ $\mu\text{g/kg}$; ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg ; wt. % (weight per cent) $\equiv$ % (mass fraction). Note 1: intervals may appear asymmetric due to rounding; IND = indeterminate.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 6 continued.

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5 % window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
4-Acid Digestion continued											
Eu, ppm	0.84	0.030	0.78	0.90	0.75	0.94	3.61%	7.22%	10.83%	0.80	0.89
Fe, wt. %	4.20	0.123	3.95	4.44	3.83	4.57	2.94%	5.88%	8.81%	3.99	4.41
Ga, ppm	19.3	0.84	17.6	21.0	16.8	21.8	4.35%	8.71%	13.06%	18.3	20.3
Gd, ppm	2.37	0.194	1.99	2.76	1.79	2.95	8.16%	16.32%	24.48%	2.25	2.49
Hf, ppm	1.52	0.081	1.36	1.69	1.28	1.77	5.29%	10.57%	15.86%	1.45	1.60
Ho, ppm	0.36	0.030	0.30	0.41	0.27	0.44	8.30%	16.59%	24.89%	0.34	0.37
In, ppm	0.032	0.003	0.026	0.038	0.023	0.041	9.06%	18.12%	27.18%	0.030	0.034
K, wt. %	1.16	0.031	1.10	1.22	1.07	1.25	2.66%	5.32%	7.97%	1.10	1.22
La, ppm	15.8	1.02	13.8	17.9	12.7	18.9	6.47%	12.93%	19.40%	15.0	16.6
Li, ppm	19.7	1.24	17.2	22.2	16.0	23.4	6.29%	12.57%	18.86%	18.7	20.7
Lu, ppm	0.14	0.010	0.12	0.16	0.11	0.17	7.13%	14.25%	21.38%	0.14	0.15
Mg, wt. %	1.65	0.041	1.57	1.73	1.53	1.77	2.48%	4.96%	7.45%	1.57	1.73
Mn, wt. %	0.067	0.002	0.062	0.072	0.060	0.074	3.61%	7.22%	10.83%	0.064	0.070
Mo, ppm	0.81	0.057	0.70	0.93	0.64	0.99	7.04%	14.09%	21.13%	0.77	0.86
Na, wt. %	3.04	0.081	2.88	3.20	2.80	3.28	2.66%	5.33%	7.99%	2.89	3.19
Nb, ppm	3.05	0.122	2.81	3.30	2.69	3.42	3.99%	7.98%	11.97%	2.90	3.20
Nd, ppm	15.6	1.15	13.3	17.9	12.1	19.0	7.36%	14.71%	22.07%	14.8	16.4
Ni, ppm	38.5	1.51	35.4	41.5	33.9	43.0	3.93%	7.87%	11.80%	36.6	40.4
P, wt. %	0.053	0.003	0.048	0.059	0.045	0.061	5.17%	10.35%	15.52%	0.051	0.056
Pb, ppm	5.40	0.390	4.62	6.17	4.23	6.56	7.23%	14.45%	21.68%	5.13	5.66
Pr, ppm	3.96	0.308	3.34	4.58	3.04	4.88	7.77%	15.54%	23.31%	3.76	4.16
Rb, ppm	33.7	3.11	27.5	39.9	24.4	43.0	9.21%	18.42%	27.63%	32.0	35.4
S, wt. %	0.050	0.000	0.049	0.051	0.048	0.051	0.95%	1.90%	2.86%	0.047	0.052
Sb, ppm	0.10	0.02	0.07	0.14	0.05	0.15	16.74%	33.49%	50.23%	0.10	0.11
Sc, ppm	14.9	0.60	13.7	16.1	13.1	16.7	4.02%	8.04%	12.06%	14.2	15.6
Sm, ppm	2.85	0.178	2.49	3.21	2.31	3.39	6.26%	12.53%	18.79%	2.71	2.99
Sn, ppm	0.60	0.030	0.54	0.66	0.51	0.69	5.07%	10.14%	15.21%	0.57	0.63
Sr, ppm	467	18	430	503	412	521	3.92%	7.83%	11.75%	443	490
Ta, ppm	0.21	0.03	0.16	0.27	0.14	0.29	11.87%	23.75%	35.62%	0.20	0.23
Tb, ppm	0.33	0.014	0.30	0.35	0.28	0.37	4.37%	8.73%	13.10%	0.31	0.34
Th, ppm	2.42	0.208	2.01	2.84	1.80	3.05	8.60%	17.19%	25.79%	2.30	2.54
Ti, wt. %	0.318	0.009	0.300	0.337	0.290	0.347	2.94%	5.89%	8.83%	0.303	0.334
Tl, ppm	0.18	0.015	0.16	0.21	0.14	0.23	7.94%	15.88%	23.82%	0.18	0.19
Tm, ppm	0.15	0.009	0.13	0.17	0.12	0.18	6.43%	12.87%	19.30%	0.14	0.15
U, ppm	0.48	0.034	0.42	0.55	0.38	0.59	7.07%	14.13%	21.20%	0.46	0.51
V, ppm	111	3	106	116	103	119	2.36%	4.72%	7.08%	105	116
W, ppm	0.60	0.059	0.49	0.72	0.43	0.78	9.70%	19.40%	29.10%	0.57	0.63
Y, ppm	10.1	0.43	9.3	11.0	8.9	11.4	4.21%	8.41%	12.62%	9.6	10.7
Yb, ppm	0.94	0.075	0.79	1.09	0.72	1.17	7.94%	15.88%	23.82%	0.90	0.99
Zn, ppm	60	3.2	53	66	50	69	5.41%	10.82%	16.23%	57	63
Zr, ppm	49.7	4.00	41.7	57.7	37.7	61.7	8.05%	16.10%	24.15%	47.2	52.2

SI unit equivalents: ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg; wt. % (weight per cent) $\equiv$ % (mass fraction).

Note 1: intervals may appear asymmetric due to rounding; IND = indeterminate.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 6 continued.

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5 % window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
Aqua Regia Digestion											
Ag, ppm	0.057	0.006	0.045	0.069	0.039	0.075	10.69%	21.37%	32.06%	0.054	0.060
Al, wt. %	1.29	0.073	1.14	1.43	1.07	1.50	5.64%	11.29%	16.93%	1.22	1.35
As, ppm	0.85	0.14	0.57	1.14	0.43	1.28	16.65%	33.30%	49.96%	0.81	0.90
B, ppm	< 10	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Ba, ppm	67	3.9	59	75	55	78	5.76%	11.53%	17.29%	63	70
Be, ppm	0.12	0.010	0.10	0.14	0.09	0.15	8.61%	17.23%	25.84%	0.11	0.13
Bi, ppm	0.066	0.007	0.052	0.080	0.045	0.087	10.51%	21.02%	31.53%	0.063	0.070
Ca, wt. %	0.905	0.123	0.660	1.150	0.537	1.272	13.54%	27.08%	40.61%	0.860	0.950
Cd, ppm	0.029	0.004	0.020	0.038	0.016	0.042	14.86%	29.71%	44.57%	0.027	0.030
Ce, ppm	26.8	1.89	23.0	30.6	21.1	32.4	7.04%	14.09%	21.13%	25.4	28.1
Co, ppm	11.8	0.37	11.0	12.5	10.7	12.9	3.14%	6.27%	9.41%	11.2	12.4
Cr, ppm	36.8	1.98	32.9	40.8	30.9	42.7	5.37%	10.74%	16.11%	35.0	38.6
Cs, ppm	0.73	0.061	0.61	0.85	0.55	0.91	8.39%	16.79%	25.18%	0.69	0.77
Cu, ppm	39.7	1.69	36.3	43.1	34.6	44.8	4.25%	8.51%	12.76%	37.7	41.7
Dy, ppm	0.96	0.14	0.69	1.24	0.55	1.37	14.23%	28.46%	42.69%	0.92	1.01
Eu, ppm	0.43	0.08	0.27	0.59	0.19	0.67	18.84%	37.67%	56.51%	0.41	0.45
Fe, wt. %	2.10	0.22	1.65	2.54	1.43	2.77	10.66%	21.33%	31.99%	1.99	2.20
Ga, ppm	4.48	0.375	3.73	5.23	3.36	5.61	8.36%	16.71%	25.07%	4.26	4.71
Gd, ppm	1.47	0.19	1.10	1.85	0.91	2.04	12.78%	25.55%	38.33%	1.40	1.55
Hf, ppm	0.19	0.02	0.15	0.23	0.13	0.26	11.06%	22.11%	33.17%	0.18	0.20
Ho, ppm	0.17	0.03	0.12	0.22	0.09	0.25	15.98%	31.96%	47.94%	0.16	0.18
In, ppm	0.009	0.003	0.004	0.015	0.001	0.017	28.77%	57.54%	86.31%	0.009	0.010
K, wt. %	0.264	0.014	0.236	0.293	0.222	0.307	5.33%	10.65%	15.98%	0.251	0.278
La, ppm	12.6	0.63	11.3	13.8	10.7	14.5	5.01%	10.01%	15.02%	12.0	13.2
Li, ppm	14.9	0.89	13.2	16.7	12.3	17.6	5.93%	11.87%	17.80%	14.2	15.7
Lu, ppm	0.053	0.005	0.044	0.062	0.039	0.066	8.58%	17.17%	25.75%	0.050	0.055
Mg, wt. %	0.808	0.039	0.731	0.885	0.693	0.924	4.76%	9.53%	14.29%	0.768	0.849
Mn, wt. %	0.030	0.001	0.028	0.032	0.027	0.033	3.34%	6.69%	10.03%	0.028	0.031
Mo, ppm	0.72	0.067	0.58	0.85	0.52	0.92	9.32%	18.63%	27.95%	0.68	0.75
Na, wt. %	0.095	0.006	0.084	0.107	0.078	0.113	6.11%	12.22%	18.33%	0.090	0.100
Nb, ppm	0.16	0.04	0.08	0.24	0.04	0.28	24.75%	49.50%	74.25%	0.15	0.17
Nd, ppm	11.5	0.95	9.6	13.4	8.7	14.4	8.21%	16.42%	24.63%	11.0	12.1
Ni, ppm	23.4	1.11	21.2	25.6	20.1	26.7	4.72%	9.44%	14.15%	22.3	24.6
P, wt. %	0.051	0.002	0.047	0.055	0.045	0.057	3.78%	7.57%	11.35%	0.049	0.054
Pb, ppm	1.50	0.25	1.00	1.99	0.76	2.24	16.51%	33.01%	49.52%	1.42	1.57
Pr, ppm	3.05	0.211	2.63	3.47	2.42	3.68	6.91%	13.82%	20.73%	2.89	3.20
Rb, ppm	13.4	0.64	12.2	14.7	11.5	15.4	4.75%	9.50%	14.25%	12.8	14.1
Re, ppm	0.000	0.000	0.000	0.001	0.000	0.001	25.31%	50.62%	75.92%	0.000	0.000
S, wt. %	0.050	0.001	0.049	0.051	0.048	0.051	1.05%	2.10%	3.14%	0.047	0.052
Sc, ppm	3.20	0.243	2.72	3.69	2.47	3.93	7.58%	15.15%	22.73%	3.04	3.36
Sm, ppm	1.82	0.164	1.49	2.15	1.33	2.31	8.99%	17.98%	26.96%	1.73	1.91
Sn, ppm	0.25	0.04	0.18	0.32	0.14	0.36	14.51%	29.02%	43.53%	0.24	0.26
Sr, ppm	45.4	5.6	34.1	56.7	28.4	62.3	12.44%	24.88%	37.32%	43.1	47.6

SI unit equivalents: ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg; wt. % (weight per cent) $\equiv$ % (mass fraction).

Note 1: intervals may appear asymmetric due to rounding; IND = indeterminate.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 6 continued.

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5 % window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
Aqua Regia Digestion continued											
Ta, ppm	< 0.01	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Te, ppm	0.020	0.007	0.006	0.034	0.000	0.040	34.14%	68.28%	102.41%	0.019	0.021
Th, ppm	2.04	0.161	1.72	2.36	1.55	2.52	7.91%	15.81%	23.72%	1.94	2.14
Ti, wt. %	0.133	0.019	0.095	0.170	0.077	0.189	14.12%	28.24%	42.35%	0.126	0.139
Tl, ppm	0.094	0.009	0.076	0.111	0.068	0.120	9.35%	18.70%	28.05%	0.089	0.099
U, ppm	0.25	0.020	0.21	0.29	0.19	0.31	7.96%	15.93%	23.89%	0.24	0.27
V, ppm	43.5	2.57	38.4	48.7	35.8	51.2	5.90%	11.80%	17.69%	41.4	45.7
W, ppm	0.19	0.03	0.13	0.24	0.11	0.26	13.87%	27.75%	41.62%	0.18	0.19
Y, ppm	4.17	0.57	3.02	5.31	2.45	5.88	13.73%	27.45%	41.18%	3.96	4.37
Yb, ppm	0.38	0.04	0.30	0.47	0.25	0.51	11.37%	22.73%	34.10%	0.36	0.40
Zn, ppm	40.6	2.15	36.3	44.9	34.1	47.0	5.29%	10.58%	15.87%	38.5	42.6
Zr, ppm	4.97	0.66	3.64	6.29	2.98	6.96	13.33%	26.66%	39.99%	4.72	5.22

SI unit equivalents: ppm (parts per million; 1×10^{-6}) $\equiv$ mg/kg; wt. % (weight per cent) $\equiv$ % (mass fraction).

Note 1: intervals may appear asymmetric due to rounding; IND = indeterminate.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

PARTICIPATING LABORATORIES

1. Actlabs, Ancaster, Ontario, Canada
2. AGAT Laboratories, Calgary, Alberta, Canada
3. ALS, Brisbane, QLD, Australia
4. ALS, Lima, Peru
5. ALS, Loughrea, Galway, Ireland
6. ALS, Malaga, WA, Australia
7. ALS, Vancouver, BC, Canada
8. American Assay Laboratories, Sparks, Nevada, USA
9. ESAN Istanbul, Istanbul, Turkey
10. Intertek, Perth, WA, Australia
11. MSALABS, Vancouver, BC, Canada
12. Ontario Geological Survey, Sudbury, Ontario, Canada
13. Paragon Geochemical Laboratories, Sparks, Nevada, USA
14. PT Geoservices Ltd, Cikarang, Jakarta Raya, Indonesia
15. PT Intertek Utama Services, Jakarta Timur, DKI Jakarta, Indonesia
16. Reminex Centre de Recherche, Marrakesh, Marrakesh-Safi, Morocco
17. SGS Canada Inc., Vancouver, BC, Canada
18. SGS Geosol Laboratorios Ltda, Vespasiano, Minas Gerais, Brazil
19. Shiva Analyticals Ltd, Bangalore North, Karnataka, India
20. Stewart Assay & Environmental Laboratories LLC, Kara-Balta, Chüy, Kyrgyzstan

PREPARER AND SUPPLIER

Certified reference material OREAS 48 is prepared, certified and supplied by:

ORE Research & Exploration Pty Ltd	Tel: +61 3 9729 0333
37A Hosie Street	Web: www.oreas.com
Bayswater North VIC 3153	Email: info@ore.com.au
AUSTRALIA	

METROLOGICAL TRACEABILITY

The interlaboratory results that underpin the certified values are metrologically traceable to the international measurement scale (SI) of mass (either as a % mass fraction or as milligrams per kilogram (mg/kg)) [14]. In line with popular use, all data within tables in this certificate are expressed as the mass fraction in either weight percent (wt.%) or parts per million (ppm) or parts per billion (ppb).

The analytical samples sent to participating laboratories were selected in a manner to be representative of the entire prepared batch of CRM. This representativeness was maintained in each submitted laboratory sample batch and ensures the user that the data is traceable from sample selection through to the analytical results. The systematic sampling method was chosen due to the low risk of overlooking repetitive effects or trends in the batch due to the way the CRM was processed. In line with ISO 17025 [8], each analytical data set received from the participating laboratories has been validated by its assayer through the inclusion of internal reference materials and QC checks during and post analysis.

The participating laboratories were chosen on the basis of their competence (from past performance in interlaboratory programs undertaken by ORE Pty Ltd) for a particular analytical method, analyte or analyte suite and sample matrix. These laboratories are accredited to ISO 17025 for Au by fire assay (Table 1). The other operationally defined measurands characterised in this certificate (Table 2) are derived from data procured mostly from ISO 17025 accredited laboratories. The certified values presented in this report are calculated from the means of accepted data following robust technical and statistical analysis as detailed in this report.

Guide ISO/TR 16476:2016 [7], section 5.3.1 describes metrological traceability in reference materials as it pertains to the transformation of the measurand. In this section it states, *“Although the determination of the property value itself can be made traceable to appropriate units through, for example, calibration of the measurement equipment used, steps like the transformation of the sample from one physical (chemical) state to another cannot. Such transformations may only be compared with a reference (when available), or among themselves. For some transformations, reference methods have been defined and may be used in certification projects to evaluate the uncertainty associated with such a transformation. In other cases, only a comparison among different laboratories using the same procedure is possible. In this case, it is impossible to demonstrate absence of method bias; therefore, the result is an operationally defined measurand (ISO 33405:2024-05, 9.2.4c) [4].”* Certification takes place on the basis of agreement among operationally defined, independent measurement results.

COMMUTABILITY

The measurements of the results that underlie the certified values contained in this report were undertaken by methods involving pre-treatment (fusion/digestion) of the sample. This served to reduce the sample to a simple and well understood form permitting calibration using simple solutions of the CRM. Due to these methods being well understood and highly effective, commutability is not an issue for this CRM. All OREAS CRMs are sourced from natural ore minerals meaning they will display similar behaviour as routine 'field' samples in the relevant measurement process. Care should be taken to ensure 'matrix matching' as close as practically achievable. The matrix and mineralisation style of the CRM is described in the 'Source Material' section and users should select appropriate CRMs matching these attributes to the field samples being analysed.

INTENDED USE

OREAS 48 is intended to cover all activities needed to produce a measurement result. This includes extraction, possible separation steps and the actual measurement process (the signal producing step). OREAS 48 may be used to calibrate the entire procedure by producing a pure substance CRM transformed into a calibration solution.

OREAS 48 is intended for the following uses:

- For the monitoring of laboratory performance in the analysis of analytes reported in Tables 1 and 2 in geological samples;
- For the verification/ validation of analytical methods for analytes reported in Tables 1 and 2;
- For the calibration of instruments used in the determination of the concentration of analytes reported in Tables 1 and 2. When a value provided in this certificate is used to calibrate a measurement process, the uncertainty associated with that value should be appropriately propagated into the user's uncertainty calculation. Users can determine an approximation of the standard uncertainty by calculating one fourth of the width of the Expanded Uncertainty interval given in this certificate (Expanded Uncertainty intervals are provided in Tables 1 and 2).

MINIMUM SAMPLE SIZE

To relate analytical determinations to the values in this certificate, the minimum mass of sample used should match the typical mass that the laboratories used in the interlaboratory (round robin) certification program. This means that different minimum sample masses should be used depending on the operationally defined methodology as follows:

- Gold by fire assay: ≥ 25 g;
- Gold by aqua regia digestion: ≥ 10 g;
- Borate/Peroxide fusion with ICP-OES and/or MS finish: ≥ 0.1 g;
- Multi-elements by 4-acid digestion with ICP-OES and/or MS finish: ≥ 0.25 g;
- Multi-elements by aqua regia digestion with ICP-OES and/or MS finish: ≥ 0.5 g.

PERIOD OF VALIDITY & STORAGE INSTRUCTIONS

The certification of OREAS 48 remains valid, within the specified measurement uncertainties, until at least June 2039, provided the CRM is handled and stored in accordance with the instructions given below. This certification is nullified if the CRM is any way changed or contaminated.

Store in a clean and cool dry place away from direct sunlight.

Long-term stability will be monitored at appropriate intervals and purchasers notified if any changes are observed. The period of validity may well be indefinite and will be reassessed prior to expiry with the aim of extending the validity if possible.

Single-use sachets

OREAS 48 is packaged in single-use laminated foil sachets. Following analysis, it is the manufacturer's expectation that any remaining material is discarded. It is the user's responsibility to prevent contamination and avoid prolonged exposure of the sample to the atmosphere prior to analysis.

Repeat-use packaging (e.g., 1 kg plastic jars)

After taking a subsample, users should replace the lid of the jar promptly and securely to prevent accidental spills and airborne contamination. OREAS 48 contains a non-hygroscopic* matrix with an indicative value for moisture provided to enable users to check for changes to stored material by determining moisture in the user's laboratory and comparing the result to the value in Table 4 in this certificate.

The stability of the CRM in regard to oxidation from the breakdown of sulphide minerals to sulphates is negligible given its low sulphur concentration (~0.05 wt.% S).

*A non-hygroscopic matrix means exposure to atmospheres significantly different, in terms of temperature and humidity, from the climate during manufacturing should have negligible impact on the precision of results. Hygroscopic moisture is the amount of adsorbed moisture (weakly held H₂O- molecules on the surface of exposed material) following exposure to the local atmosphere. Usually, equilibration of material to the local atmosphere will only occur if the material is spread into a thin (~2mm thick) layer and left exposed for a period of 2 hours.

INSTRUCTIONS FOR HANDLING & CORRECT USE

Pre-homogenisation of the CRM prior to subsampling and analysis is not necessary as there is no particle segregation under transport [12].

Fine powders pose a risk to eyes and lungs and therefore standard precautions including the use of safety glasses and dust masks are advised.

QC monitoring using multiples of the Standard Deviation (SD)

In the application of SD's in monitoring performance it is important to note that not all laboratories function at the same level of proficiency and that different methods in use at a particular laboratory have differing levels of precision. Each laboratory has its own inherent SD (for a specific concentration level and analyte-method pair) based on the analytical process and this SD is not directly related to the round robin program.

The majority of data generated in the round robin program was produced by a selection of world class laboratories. The SD's thus generated are more constrained than those that would be produced across a randomly selected group of laboratories. To produce more generally achievable SD's the 'pooled' SD's provided in this report include interlaboratory bias. This 'one size fits all' approach may require revision at the discretion of the QC manager concerned following careful scrutiny of QC control charts.

The performance gates shown in Table 6 are intended only to be used as a preliminary guide as to what a laboratory may be able to achieve. Over a period of time monitoring your own laboratory's data for this CRM, SD's should be calculated directly from your own laboratory's process. This will enable you to establish more specific performance gates that are fit for purpose for your application as well as the ability to monitor bias. If your long-term trend analysis shows an average value that is within the 95 % expanded uncertainty then generally there is no cause for concern in regard to bias.

For use with the aqua regia digestion method

It is important to note that in the analytical industry there is no standardisation of the aqua regia digestion process. This method is a partial empirical digest and differences in recoveries for various analytes are commonplace. These are caused by variations in the digest conditions and can include the ratio of nitric to hydrochloric acids, acid strength, temperatures, leach times and secondary digestions. Recoveries for sulphide-hosted base metal sulphides approach total values, however, other analytes, particularly the lithophile elements, show greater sensitivity to method parameters. This can result in lack of consensus in an inter-laboratory certification program for these elements.

The approach applied here is to report certified values in those instances where reasonable agreement exists amongst a majority of participating laboratories. The results of specific laboratories may differ significantly from the certified values, but will, nonetheless, be valid and reproducible in the context of the specifics of the aqua regia method in use. Users of this reference material should, therefore, be mindful of this limitation when applying the certified values in a quality control program.

LEGAL NOTICE

Ore Research & Exploration Pty Ltd has prepared and statistically evaluated the property values of this reference material to the best of its ability. The Purchaser by receipt hereof releases and indemnifies Ore Research & Exploration Pty Ltd from and against all liability and costs arising from the use of this material and information.

**© COPYRIGHT Ore Research & Exploration Pty Ltd.
Unauthorised copying, reproduction, storage or dissemination is prohibited.**

QMS CERTIFICATION

ORE Pty Ltd is accredited for compliance with ISO 17034:2016 (Accreditation number 20483).



ORE Pty Ltd is ISO 9001:2015 certified by Lloyd's Register Quality Assurance Ltd for its quality management system including development, manufacturing, certification and supply of CRMs.



DOCUMENT HISTORY

Revision No.	Date	Changes applied
0	6 th May, 2025	First publication.

CERTIFYING OFFICER

Craig Hamlyn (B.Sc. Hons - Geology), Technical Manager - ORE P/L

REFERENCES

- [1] Govett, G.J.S. (1983). Handbook of Exploration Geochemistry, Volume 2: Statistics and Data Analysis in Geochemical Prospecting (Variations of accuracy and precision).
- [2] ISO Guide 30:2015. Terms and definitions used in connection with reference materials.
- [3] ISO 33401:2024-01. Reference materials – Contents of certificates, labels and accompanying documentation.
- [4] ISO 33405:2024-05. Reference materials – Approaches for characterization and assessment of homogeneity and stability.
- [5] ISO Guide 98-3:2008. Guide to the expression of uncertainty in measurement (GUM:1995).
- [6] ISO 16269:2014. Statistical interpretation of data – Part 6: Determination of statistical tolerance intervals.
- [7] ISO/TR 16476:2016, Reference Materials – Establishing and expressing metrological traceability of quantity values assigned to reference materials.
- [8] ISO 17025:2017, General requirements for the competence of testing and calibration laboratories.
- [9] ISO 17034:2016. General requirements for the competence of reference material producers.
- [10] Munsell Rock Color Book (2014). Rock-Color Chart Committee, Geological Society of America (GSA), Minnesota (USA).
- [11] OREAS-BUP-70-09-11: Statistical Analysis - OREAS Evaluation Method.
- [12] OREAS-TN-04-1498: Stability under transport; an experimental study of OREAS CRMs.
- [13] OREAS-TN-05-1674: Long-term storage stability; an experimental study of OREAS CRMs.
- [14] Thompson, A.; Taylor, B.N. (2008); Guide for the Use of the International System of Units (SI); NIST Special Publication 811; U.S. Government Printing Office: Washington, DC; available at: <https://physics.nist.gov/cuu/pdf/sp811.pdf> (accessed Nov 2021).
- [15] Van der Veen A.M.H. et al. (2001). Uncertainty calculations in the certification of reference materials, Accred Qual Assur 6: 290-294.